

TABLA NMP PARA AGUAS POTABLES

Table 9221:2. MPN Index and 95% Confidence Limits for All Combinations of Positive and Negative Results When Five 20-mL Portions Are Used

No. of Tubes Giving Positive Reaction Out of 5 (20 mL Each)	MPN Index/ 100 mL	95% Confidence Limits (Exact)	
		Lower	Upper
0	<1.1	—	3.5
1	1.1	0.051	5.4
2	2.6	0.40	8.4
3	4.6	1.0	13
4	8.0	2.1	23
5	>8.0	3.4	—